
Midmark M11 - v1.0.0
Total Cycles: 3549

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2023 11:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.7 F 0.0
2:00 195.9 F 3.6
4:00 208.6 F 8.5
6:00 223.8 F 9.7
8:00 228.5 F 9.5
10:00 234.2 F 8.8
12:00 244.1 F 12.9
14:00 242.5 F 11.5
16:00 255.5 F 18.2
18:00 266.9 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.2 F 28.7
2:00 271.7 F 28.6
3:00 272.1 F 28.7
4:00 271.4 F 28.4

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 224.8 F 0.0
10:00 245.8 F 0.0
15:00 242.3 F 0.0
20:00 227.5 F 0.0
25:00 223.4 F 0.0
30:00 218.7 F 0.0

Drying Complete

FILLING: 1:50
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

07/19/2023 12:29 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____