

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000435  
9:30 28/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:13  
STERILIZING 4:12  
136.5 C 232kPa 4:12  
136.4 C 232kPa 4:42  
136.2 C 232kPa 5:12  
136.3 C 232kPa 5:42  
136.4 C 233kPa 6:12  
136.4 C 233kPa 6:42  
136.4 C 233kPa 7:12  
136.4 C 233kPa 7:42  
136.4 C 233kPa 7:42  
VENTING 7:42  
AIR DRYING 8:38  
CYCLE COMPLETE 8:40  
-----

Digital Signature #  
1F10EAC34A49CF29  
-----