
Midmark M11 - v1.0.5
Total Cycles: 341

Name: RITTER2
Sterilizer ID: V2313228
09/08/2021 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 161.6 F 3.1
4:00 174.7 F 7.7
6:00 201.8 F 13.3
8:00 218.0 F 8.7
10:00 231.4 F 10.4
12:00 238.3 F 12.4
14:00 240.3 F 11.4
16:00 253.4 F 18.1
18:00 264.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 271.9 F 29.6
2:00 272.2 F 30.0
3:00 271.5 F 29.5
4:00 271.6 F 29.7

Min 271.2 F 29.2
Max 272.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 201.6 F 0.0
10:00 209.9 F 0.0
15:00 213.6 F 0.0
20:00 214.7 F 0.0
25:00 215.3 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

09/08/2021 01:22 PM
ACCEPTED: X
REJECTED:
Operator: U01