
Midmark M11 - v1.0.5
Total Cycles: 4267

Name: RITTER2
Sterilizer ID: V2313228
01/15/2025 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.8 F 0.0
2:00 89.6 F 0.0
4:00 97.6 F 0.0
6:00 128.5 F 0.2
8:00 160.7 F 0.5
10:00 185.7 F 2.3
12:00 201.9 F 5.6
14:00 218.4 F 9.5
16:00 231.5 F 11.4
18:00 234.0 F 9.5
20:00 239.2 F 10.5
22:00 243.8 F 12.6
24:00 242.0 F 11.5
26:00 254.1 F 17.7
28:00 264.8 F 24.1
30:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.3 F 29.2
2:00 272.7 F 29.5
3:00 272.5 F 29.8
4:00 271.9 F 29.2

Min 271.9 F 29.1
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 220.1 F 0.0
10:00 186.4 F 0.0
15:00 184.6 F 0.0
20:00 181.9 F 0.0
25:00 182.1 F 0.0
30:00 185.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:09:19
CYCLE COMPLETE

01/15/2025 04:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____