
Midmark M11 - v1.0.0
Total Cycles: 1562

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2022 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.4 F 0.0
2:00 186.2 F 4.7
4:00 199.3 F 10.0
6:00 215.5 F 5.4
8:00 231.4 F 10.5
10:00 231.9 F 7.1
12:00 244.1 F 12.2
14:00 245.4 F 12.7
16:00 258.5 F 19.7
18:00 270.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.4 F 28.4
2:00 271.6 F 28.3
3:00 272.4 F 28.5
4:00 271.3 F 28.0

Min 271.3 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 211.8 F 0.0
10:00 230.1 F 0.0
15:00 226.6 F 0.0
20:00 222.6 F 0.0
25:00 219.7 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

01/25/2022 03:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____