
Midmark M11 - v1.0.5
Total Cycles: 423

Name: RITTER2
Sterilizer ID: V2313228
10/06/2021 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.8 F 0.0
2:00 112.4 F 1.0
4:00 118.3 F 3.4
6:00 148.2 F 6.8
8:00 182.7 F 11.4
10:00 227.3 F 6.0
12:00 235.2 F 10.7
14:00 231.1 F 7.2
16:00 241.5 F 12.1
18:00 238.7 F 9.8
20:00 250.8 F 15.7
22:00 261.2 F 21.7
24:00 271.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 29.6
1:00 272.1 F 29.9
2:00 272.3 F 29.9
3:00 271.4 F 30.0
4:00 272.2 F 30.4

Min 270.7 F 29.5
Max 273.6 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 219.9 F 0.0
10:00 237.4 F 0.0
15:00 227.4 F 0.0
20:00 224.9 F 0.0
25:00 228.9 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

10/06/2021 04:07 PM
ACCEPTED: X
REJECTED:
Operator: U01