
Midmark M11 - v1.0.5
Total Cycles: 4206

Name: RITTER2
Sterilizer ID: V2313228
12/12/2024 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 77.3 F 0.0
4:00 102.4 F 0.0
6:00 108.0 F 0.1
8:00 151.2 F 0.7
10:00 180.2 F 3.8
12:00 194.2 F 7.4
14:00 207.3 F 11.1
16:00 216.0 F 5.0
18:00 228.0 F 8.5
20:00 238.4 F 12.8
22:00 237.5 F 10.1
24:00 236.8 F 8.9
26:00 248.7 F 15.0
28:00 260.3 F 21.5
30:00 270.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.4 F 29.1
2:00 272.6 F 29.6
3:00 271.6 F 28.7
4:00 272.7 F 29.4

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.2 F 0.0
5:00 212.5 F 0.0
10:00 217.5 F 0.0
15:00 196.8 F 0.0
20:00 195.8 F 0.0
25:00 200.1 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:42
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:08:49
CYCLE COMPLETE

12/12/2024 11:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____