
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001229
11:11 26/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:40
STERILIZING 3:46
136.0 C 220kPa 3:46
135.9 C 219kPa 4:16
136.4 C 224kPa 4:46
136.2 C 223kPa 5:16
136.4 C 225kPa 5:46
136.4 C 223kPa 6:16
136.4 C 224kPa 6:46
136.4 C 224kPa 7:16
136.4 C 224kPa 7:16
VENTING 7:16
AIR DRYING 8:09
CYCLE COMPLETE 23:09

Digital Signature #
22D21F1ECECF6FDB
