
Midmark M11 - v1.0.0
Total Cycles: 1449

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2021 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 196.2 F 2.7
4:00 206.7 F 6.5
6:00 219.1 F 11.6
8:00 226.6 F 6.1
10:00 236.3 F 11.0
12:00 236.8 F 7.7
14:00 244.4 F 11.9
16:00 244.3 F 12.6
18:00 256.7 F 19.2
20:00 267.7 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.4
2:00 271.9 F 28.8
3:00 271.4 F 28.9
4:00 272.5 F 29.4

Min 271.2 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 229.1 F 0.0
10:00 244.4 F 0.0
15:00 230.0 F 0.0
20:00 227.1 F 0.0
25:00 229.5 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

12/13/2021 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____