

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000414  
15:15 26/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:09  
STERILIZING 4:12  
136.0 C 229kPa 4:12  
136.4 C 232kPa 4:42  
136.4 C 233kPa 5:12  
136.4 C 233kPa 5:42  
136.3 C 232kPa 6:12  
136.3 C 233kPa 6:42  
136.4 C 233kPa 7:12  
136.4 C 233kPa 7:42  
136.4 C 233kPa 7:42  
VENTING 7:42  
AIR DRYING 8:36  
CYCLE COMPLETE 8:38

-----  
Digital Signature #  
D777850E4A05DE82  
-----