
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002628
7:48 28/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:30
STERILIZING 3:39
136.5 C 225kPa 3:39
136.1 C 222kPa 4:09
136.3 C 223kPa 4:39
136.3 C 224kPa 5:09
136.2 C 223kPa 5:39
136.3 C 224kPa 6:09
136.4 C 225kPa 6:39
136.4 C 225kPa 7:09
136.4 C 225kPa 7:09
VENTING 7:09
AIR DRYING 8:06
CYCLE COMPLETE 38:06

Digital Signature #
B10E4D0B511AEAF9
