

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001081  
15:15 17/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:28  
PRESSURIZING 3:30  
STERILIZING 4:35  
136.0 C 220kPa 4:35  
136.4 C 224kPa 5:05  
136.2 C 223kPa 5:35  
136.2 C 223kPa 6:05  
136.4 C 224kPa 6:35  
136.2 C 223kPa 7:05  
136.3 C 224kPa 7:35  
136.4 C 225kPa 8:05  
136.4 C 225kPa 8:05  
VENTING 8:05  
AIR DRYING 8:59  
CYCLE COMPLETE 23:59  
-----

Digital Signature #  
4915CAAA029BB86E  
-----