

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4246

Name: RITTER2  
Sterilizer ID: V2313228  
01/08/2025 09:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.7 F 0.0  
2:00 81.4 F 0.0  
4:00 119.4 F 0.0  
6:00 110.1 F 0.1  
8:00 152.6 F 0.6  
10:00 175.5 F 3.3  
12:00 193.1 F 6.6  
14:00 207.3 F 10.4  
16:00 219.4 F 7.2  
18:00 227.9 F 8.0  
20:00 239.5 F 12.3  
22:00 235.9 F 9.2  
24:00 240.2 F 10.4  
26:00 246.6 F 13.7  
28:00 258.2 F 20.0  
30:00 268.6 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 271.8 F 28.6  
2:00 272.0 F 28.8  
3:00 272.5 F 29.0  
4:00 272.8 F 29.3

Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 238.4 F 0.0  
10:00 221.0 F 0.0  
15:00 215.1 F 0.0  
20:00 221.1 F 0.0  
25:00 229.1 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:05  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

01/08/2025 10:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_