
Midmark M11 - v1.0.0
Total Cycles: 2318

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2022 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.2 F 0.0
2:00 188.2 F 3.7
4:00 206.8 F 8.5
6:00 223.1 F 7.8
8:00 233.8 F 9.3
10:00 237.5 F 9.5
12:00 242.8 F 12.1
14:00 240.2 F 10.4
16:00 253.8 F 17.2
18:00 264.9 F 24.0
20:00 271.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 271.7 F 28.9
2:00 271.5 F 28.8
3:00 271.8 F 28.8
4:00 272.1 F 28.9

Min 271.5 F 28.7
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 238.7 F 0.0
10:00 227.0 F 0.0
15:00 226.5 F 0.0
20:00 223.8 F 0.0
25:00 227.1 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

09/13/2022 11:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____