
Midmark M11 - v1.0.0
Total Cycles: 2426

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2022 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.7 F 0.0
2:00 150.8 F 1.4
4:00 159.2 F 4.2
6:00 176.1 F 8.0
8:00 194.6 F 12.2
10:00 213.0 F 4.3
12:00 224.7 F 8.5
14:00 235.5 F 12.9
16:00 234.9 F 9.2
18:00 239.8 F 10.3
20:00 244.3 F 12.6
22:00 255.9 F 18.7
24:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 271.6 F 28.6
2:00 272.1 F 29.1
3:00 271.5 F 28.8
4:00 271.6 F 28.6

Min 271.3 F 28.4
Max 272.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 222.7 F 0.0
10:00 238.6 F 0.0
15:00 233.0 F 0.0
20:00 226.9 F 0.0
25:00 223.0 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

10/13/2022 09:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____