

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001411  
8:07 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:28  
PRESSURIZING 6:04  
STERILIZING 7:35  
136.4 C 221kPa 7:35  
136.1 C 220kPa 8:05  
136.3 C 221kPa 8:35  
136.4 C 223kPa 9:05  
136.5 C 222kPa 9:35  
136.5 C 222kPa 10:05  
136.1 C 220kPa 10:35  
136.5 C 219kPa 11:05  
VENTING 11:05  
AIR DRYING 12:06  
CYCLE COMPLETE 12:16  
-----

Digital Signature #  
A61BB901FB766C23  
-----