
Midmark M11 - v1.0.0
Total Cycles: 2909

Name: RITTER 1
Sterilizer ID: V2178198
02/15/2023 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.7 F 0.0
2:00 166.8 F 3.8
4:00 183.3 F 7.9
6:00 209.1 F 12.8
8:00 217.9 F 6.7
10:00 230.9 F 11.5
12:00 232.0 F 8.1
14:00 241.5 F 11.7
16:00 243.9 F 12.5
18:00 256.1 F 18.9
20:00 267.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.4 F 29.3
2:00 271.5 F 28.6
3:00 272.2 F 29.5
4:00 272.0 F 29.0

Min 271.3 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 211.8 F 0.0
10:00 234.0 F 0.0
15:00 236.7 F 0.0
20:00 229.0 F 0.0
25:00 230.1 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

02/15/2023 03:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____