

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001858  
9:03 16/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:09  
PRESSURIZING 3:15  
STERILIZING 4:24  
136.4 C 224kPa 4:24  
136.4 C 225kPa 4:54  
136.3 C 224kPa 5:24  
136.4 C 225kPa 5:54  
136.4 C 225kPa 6:24  
136.4 C 224kPa 6:54  
136.2 C 223kPa 7:24  
136.4 C 225kPa 7:54  
136.4 C 225kPa 7:54  
VENTING 7:54  
AIR DRYING 8:51  
CYCLE COMPLETE 8:58

-----  
Digital Signature #  
325413063A8BDC0A  
-----