
Midmark M11 - v1.0.0
Total Cycles: 2413

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2022 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 78.6 F 0.0
4:00 92.5 F 0.0
6:00 125.5 F 1.4
8:00 154.5 F 3.6
10:00 177.5 F 6.7
12:00 196.5 F 10.5
14:00 212.7 F 5.6
16:00 223.2 F 6.6
18:00 233.8 F 10.7
20:00 235.0 F 7.6
22:00 243.1 F 12.2
24:00 240.3 F 10.1
26:00 253.0 F 16.6
28:00 264.2 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 271.9 F 28.3
2:00 272.4 F 28.7
3:00 272.8 F 28.7
4:00 272.8 F 28.7

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 236.8 F 0.0
5:00 223.0 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 29:55
STERILIZING: 4:00
VENTING: 2:43
DRYING: 5:50
TOTAL CYCLE: 00:43:19

10/11/2022 09:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____