
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002190
8:58 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:01
STERILIZING 3:02
136.0 C 220kPa 3:02
136.4 C 226kPa 3:32
136.4 C 225kPa 4:02
136.2 C 224kPa 4:32
136.4 C 226kPa 5:02
136.5 C 223kPa 5:32
136.4 C 225kPa 6:02
136.2 C 223kPa 6:32
136.2 C 223kPa 6:32
VENTING 6:32
AIR DRYING 7:21
CYCLE COMPLETE 19:40

Digital Signature #
5CDAD7BCBAC2C2FB
