

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 004503  
8:20 06/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:38  
STERILIZING 3:41  
136.3 C 220kPa 3:41  
135.9 C 220kPa 4:11  
136.4 C 222kPa 4:41  
136.4 C 222kPa 5:11  
136.3 C 223kPa 5:41  
135.7 C 219kPa 6:11  
136.4 C 222kPa 6:41  
136.2 C 219kPa 7:11  
136.2 C 219kPa 7:11  
VENTING 7:11  
AIR DRYING 8:00  
CYCLE COMPLETE 8:03

-----  
Digital Signature #  
B41D78FB80692EB6  
-----