

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1242

Name: RITTER 1  
Sterilizer ID: V2178198  
10/07/2021 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.3 F 0.0  
2:00 162.3 F 2.8  
4:00 182.9 F 6.2  
6:00 209.2 F 11.0  
8:00 219.5 F 5.2  
10:00 235.4 F 10.3  
12:00 233.4 F 7.8  
14:00 245.4 F 13.3  
16:00 243.8 F 11.8  
18:00 256.6 F 18.5  
20:00 267.9 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.5 F 28.5  
2:00 271.6 F 28.1  
3:00 272.6 F 28.6  
4:00 271.6 F 27.9

Min 271.5 F 27.9  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.1 F 0.0  
5:00 231.1 F 0.0  
10:00 223.2 F 0.0  
15:00 226.4 F 0.0  
20:00 224.9 F 0.0  
25:00 210.9 F 0.0  
30:00 205.7 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

10/07/2021 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_