
Midmark M11 - v1.0.0
Total Cycles: 2999

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2023 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 191.6 F 3.7
4:00 214.4 F 9.1
6:00 226.6 F 7.8
8:00 236.8 F 11.0
10:00 236.3 F 8.8
12:00 239.9 F 9.6
14:00 249.2 F 14.9
16:00 261.8 F 22.0
18:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.8 F 29.2
4:00 272.0 F 28.6

Min 271.4 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 232.0 F 0.0
10:00 238.7 F 0.0
15:00 231.8 F 0.0
20:00 229.7 F 0.0
25:00 226.6 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

03/13/2023 04:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____