

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1918

Name: RITTER 1  
Sterilizer ID: V2178198  
05/09/2022 08:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.3 F 0.0  
2:00 124.3 F 0.0  
4:00 138.4 F 1.4  
6:00 166.7 F 4.0  
8:00 190.8 F 8.1  
10:00 219.1 F 13.0  
12:00 220.4 F 6.9  
14:00 232.0 F 11.5  
16:00 231.9 F 7.7  
18:00 244.5 F 13.4  
20:00 242.5 F 11.3  
22:00 254.5 F 17.5  
24:00 265.1 F 23.9  
26:00 271.2 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.5 F 29.1  
2:00 272.7 F 28.7  
3:00 271.4 F 28.1  
4:00 271.8 F 28.6

Min 271.2 F 28.1  
Max 273.1 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 218.0 F 0.0  
10:00 197.6 F 0.0  
15:00 197.1 F 0.0  
20:00 195.7 F 0.0  
25:00 229.1 F 0.0  
30:00 231.3 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 26:01  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:03:54  
CYCLE COMPLETE

05/09/2022 10:17 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02