
Midmark M11 - v1.0.0
Total Cycles: 1842

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2022 12:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 160.4 F 3.2
4:00 182.1 F 7.6
6:00 221.9 F 13.0
8:00 223.1 F 7.2
10:00 235.8 F 12.2
12:00 235.0 F 8.9
14:00 239.9 F 10.1
16:00 246.0 F 13.5
18:00 258.4 F 20.2
20:00 269.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.4 F 29.3
2:00 272.4 F 29.0
3:00 271.3 F 28.8
4:00 272.7 F 29.4

Min 271.1 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 231.4 F 0.0
10:00 223.3 F 0.0
15:00 225.3 F 0.0
20:00 228.0 F 0.0
25:00 230.7 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

04/19/2022 01:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____