

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4644

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2024 12:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.8 F 0.0  
2:00 154.7 F 2.6  
4:00 169.6 F 6.0  
6:00 192.5 F 10.1  
8:00 216.9 F 7.2  
10:00 222.0 F 8.5  
12:00 232.7 F 13.4  
14:00 233.3 F 9.7  
16:00 236.4 F 9.2  
18:00 246.3 F 13.5  
20:00 257.8 F 19.8  
22:00 268.4 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.1 F 28.9  
2:00 272.5 F 29.2  
3:00 272.2 F 29.3  
4:00 271.7 F 28.8

Min 271.5 F 28.7  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 224.2 F 0.0  
10:00 219.3 F 0.0  
15:00 215.4 F 0.0  
20:00 212.5 F 0.0  
25:00 212.8 F 0.0  
30:00 214.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:11  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:01:50  
CYCLE COMPLETE

06/14/2024 01:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_