

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1944

Name: RITTER 1  
Sterilizer ID: V2178198  
05/18/2022 10:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.5 F 0.0  
2:00 156.4 F 2.9  
4:00 166.9 F 6.4  
6:00 191.3 F 10.7  
8:00 216.4 F 3.6  
10:00 225.7 F 8.7  
12:00 237.0 F 12.0  
14:00 237.4 F 10.4  
16:00 235.3 F 7.9  
18:00 248.8 F 14.6  
20:00 260.6 F 21.2  
22:00 271.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.2 F 28.6  
2:00 272.1 F 29.0  
3:00 271.6 F 28.3  
4:00 272.5 F 29.0

Min 271.5 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 210.1 F 0.0  
10:00 232.1 F 0.0  
15:00 230.3 F 0.0  
20:00 229.3 F 0.0  
25:00 225.2 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

05/18/2022 11:45 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_