
Midmark M11 - v1.0.5
Total Cycles: 1142

Name: RITTER2
Sterilizer ID: V2313228
05/31/2022 08:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.7 F 0.0
2:00 162.9 F 1.4
4:00 188.8 F 5.1
6:00 200.2 F 10.2
8:00 217.9 F 4.8
10:00 226.1 F 9.5
12:00 235.4 F 9.2
14:00 240.8 F 12.4
16:00 239.9 F 11.1
18:00 252.9 F 17.6
20:00 264.1 F 24.4
22:00 271.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.1 F 29.8
2:00 272.6 F 30.1
3:00 272.0 F 30.3
4:00 271.9 F 29.7

Min 271.7 F 29.6
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 204.7 F 0.0
10:00 220.2 F 0.0
15:00 231.4 F 0.0
20:00 223.7 F 0.0
25:00 216.8 F 0.0
30:00 233.8 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:06
CYCLE COMPLETE

05/31/2022 09:04 AM
ACCEPTED: X
REJECTED:
Operator: U02