
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003951
12:23 06/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:39
STERILIZING 3:43
136.4 C 224kPa 3:43
136.1 C 223kPa 4:13
136.0 C 222kPa 4:43
136.4 C 222kPa 5:13
136.4 C 223kPa 5:43
136.4 C 224kPa 6:13
136.3 C 223kPa 6:43
136.4 C 225kPa 7:13
136.4 C 225kPa 7:13
VENTING 7:13
AIR DRYING 7:59
CYCLE COMPLETE 12:15

Digital Signature #
10B0DAB0141D4FE1
