
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000844
10:45 06/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:54
STERILIZING 4:10
136.0 C 228kPa 4:10
136.0 C 229kPa 4:40
136.3 C 232kPa 5:10
135.9 C 233kPa 5:40
136.2 C 231kPa 6:10
136.2 C 231kPa 6:40
136.3 C 231kPa 7:10
136.4 C 231kPa 7:40
136.4 C 231kPa 7:40
VENTING 7:40
AIR DRYING 8:32
CYCLE COMPLETE 23:32

Digital Signature #
6A289BCA1633B5F8
