
Midmark M11 - v1.0.5
Total Cycles: 3886

Name: RITTER2
Sterilizer ID: V2313228
08/22/2024 04:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.2 F 0.0
2:00 170.3 F 2.7
4:00 188.0 F 7.4
6:00 210.6 F 13.2
8:00 224.0 F 7.5
10:00 239.5 F 12.9
12:00 238.9 F 10.7
14:00 237.8 F 9.7
16:00 252.8 F 16.9
18:00 265.1 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.5 F 29.1
2:00 272.3 F 29.3
3:00 272.1 F 28.8
4:00 272.8 F 29.5

Min 271.7 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 222.6 F 0.0
10:00 205.4 F 0.0
15:00 204.0 F 0.0
20:00 201.9 F 0.0
25:00 199.0 F 0.0
30:00 196.3 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

08/22/2024 05:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____