
Midmark M11 - v1.0.0
Total Cycles: 2189

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2022 03:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 189.5 F 4.4
4:00 208.2 F 9.3
6:00 221.4 F 6.6
8:00 231.0 F 9.1
10:00 238.8 F 10.3
12:00 241.0 F 11.7
14:00 239.1 F 10.1
16:00 252.4 F 16.7
18:00 263.8 F 23.4
20:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.3
2:00 272.0 F 28.6
3:00 272.0 F 29.0
4:00 272.7 F 29.1

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 235.6 F 0.0
10:00 200.3 F 0.0
15:00 194.8 F 0.0
20:00 193.0 F 0.0
25:00 192.0 F 0.0
30:00 193.1 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

07/27/2022 04:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____