

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1264

Name: RITTER2  
Sterilizer ID: V2313228  
07/07/2022 06:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.2 F 0.0  
2:00 74.4 F 0.0  
4:00 79.8 F 0.0  
6:00 109.2 F 0.0  
8:00 143.4 F 0.2  
10:00 175.7 F 0.8  
12:00 196.7 F 4.4  
14:00 211.5 F 7.7  
16:00 223.5 F 11.3  
18:00 226.5 F 6.7  
20:00 236.0 F 11.3  
22:00 234.9 F 8.7  
24:00 241.3 F 10.5  
26:00 246.6 F 13.7  
28:00 258.0 F 19.9  
30:00 268.2 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.4 F 29.3  
2:00 272.5 F 29.8  
3:00 271.6 F 29.0  
4:00 272.6 F 29.6

Min 271.5 F 29.0  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 224.0 F 0.0  
10:00 217.0 F 0.0  
15:00 211.0 F 0.0  
20:00 213.9 F 0.0  
25:00 210.3 F 0.0  
30:00 197.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 32:00  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:10:03  
CYCLE COMPLETE

07/07/2022 08:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_