
Midmark M11 - v1.0.0
Total Cycles: 2372

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2022 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.9 F 0.0
2:00 181.6 F 4.4
4:00 196.6 F 9.2
6:00 210.6 F 7.9
8:00 225.0 F 8.9
10:00 234.9 F 10.4
12:00 239.9 F 11.5
14:00 238.1 F 9.9
16:00 251.8 F 16.5
18:00 263.8 F 23.6
20:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.5
2:00 272.0 F 28.9
3:00 271.3 F 28.8
4:00 272.5 F 29.5

Min 271.2 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 234.7 F 0.0
10:00 236.8 F 0.0
15:00 230.7 F 0.0
20:00 231.0 F 0.0
25:00 228.7 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

09/26/2022 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____