
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001839
12:19 22/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:39
136.4 C 221kPa 3:39
136.4 C 225kPa 4:09
136.4 C 225kPa 4:39
136.1 C 222kPa 5:09
136.4 C 224kPa 5:39
136.4 C 225kPa 6:09
136.2 C 221kPa 6:39
136.2 C 219kPa 7:09
136.2 C 219kPa 7:09
VENTING 7:09
AIR DRYING 8:03
CYCLE COMPLETE 8:09

Digital Signature #
1D21D2C7FF15EAB0
