
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000156
13:44 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:55
STERILIZING 2:55
136.4 C 220kPa 2:55
136.4 C 224kPa 3:25
136.2 C 221kPa 3:55
136.2 C 223kPa 4:25
136.3 C 223kPa 4:55
136.4 C 223kPa 5:25
136.2 C 220kPa 5:55
136.2 C 223kPa 6:25
136.2 C 223kPa 6:25
VENTING 6:26
AIR DRYING 7:17
CYCLE COMPLETE 7:22

Digital Signature #
67FF3C3BDB9CAA57
