

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4073

Name: RITTER2  
Sterilizer ID: V2313228  
11/01/2024 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.2 F 0.0  
2:00 72.0 F 0.0  
4:00 74.2 F 0.0  
6:00 100.8 F 0.1  
8:00 136.6 F 0.5  
10:00 163.0 F 2.4  
12:00 187.6 F 5.4  
14:00 205.1 F 9.0  
16:00 216.2 F 12.7  
18:00 228.3 F 6.9  
20:00 237.5 F 11.0  
22:00 232.2 F 7.4  
24:00 244.5 F 13.1  
26:00 241.9 F 11.7  
28:00 254.3 F 17.9  
30:00 265.0 F 24.4  
32:00 271.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.8 F 29.8  
2:00 271.6 F 29.5  
3:00 272.9 F 29.7  
4:00 271.6 F 29.5  
  
Min 271.5 F 29.1  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 170.2 F 0.0  
5:00 240.8 F 0.0  
10:00 222.8 F 0.0  
15:00 227.6 F 0.0  
20:00 217.0 F 0.0  
25:00 200.5 F 0.0  
30:00 179.4 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:10:60  
CYCLE COMPLETE

11/01/2024 10:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_