
Midmark M11 - v1.0.0
Total Cycles: 4898

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2024 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 169.2 F 0.2
4:00 181.2 F 0.6
6:00 207.5 F 3.8
8:00 224.7 F 8.0
10:00 235.4 F 12.4
12:00 232.9 F 7.8
14:00 243.9 F 13.1
16:00 241.6 F 10.9
18:00 237.8 F 9.1
20:00 251.4 F 15.6
22:00 262.3 F 21.9
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 271.8 F 28.4
2:00 272.5 F 28.8
3:00 272.5 F 28.6
4:00 271.9 F 28.3

Min 271.1 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 235.9 F 0.0
10:00 198.3 F 0.0
15:00 201.6 F 0.0
20:00 180.3 F 0.0
25:00 182.8 F 0.0
30:00 182.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:43
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

09/12/2024 11:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____