
Midmark M11 - v1.0.0
Total Cycles: 886

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2021 04:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 163.6 F 2.2
4:00 187.5 F 6.0
6:00 206.9 F 10.3
8:00 218.2 F 6.4
10:00 227.7 F 9.2
12:00 237.5 F 10.2
14:00 241.5 F 11.9
16:00 240.8 F 10.6
18:00 253.8 F 17.3
20:00 265.4 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.5 F 28.3
3:00 272.5 F 28.9
4:00 271.7 F 28.2

Min 271.5 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 213.5 F 0.0
10:00 218.7 F 0.0
15:00 225.1 F 0.0
20:00 229.6 F 0.0
25:00 232.0 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

06/10/2021 05:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____