
Midmark M11 - v1.0.0
Total Cycles: 5109

Name: RITTER 1
Sterilizer ID: V2178198
11/19/2024 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 150.9 F 0.0
4:00 162.9 F 0.3
6:00 188.9 F 0.8
8:00 205.0 F 3.9
10:00 219.0 F 7.5
12:00 229.7 F 11.5
14:00 228.3 F 6.9
16:00 240.5 F 12.1
18:00 238.0 F 9.9
20:00 238.2 F 9.1
22:00 249.3 F 14.8
24:00 260.5 F 21.0
26:00 270.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.9
2:00 272.4 F 28.7
3:00 271.9 F 28.3
4:00 271.4 F 28.2

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 212.4 F 0.0
10:00 214.3 F 0.0
15:00 222.8 F 0.0
20:00 226.1 F 0.0
25:00 215.9 F 0.0
30:00 217.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:53
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:05:32
CYCLE COMPLETE

11/19/2024 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____