

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002511  
13:53 11/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:19  
STERILIZING 4:32  
136.4 C 220kPa 4:32  
136.2 C 222kPa 5:02  
136.4 C 225kPa 5:32  
136.4 C 225kPa 6:02  
136.4 C 225kPa 6:32  
136.2 C 224kPa 7:02  
136.4 C 225kPa 7:32  
136.4 C 223kPa 8:02  
136.4 C 223kPa 8:02  
VENTING 8:02  
AIR DRYING 9:02  
CYCLE COMPLETE 9:12

-----  
Digital Signature #  
444A5010138930DA  
-----