
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001915
8:11 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:02
STERILIZING 3:08
136.4 C 223kPa 3:08
136.2 C 223kPa 3:38
136.4 C 222kPa 4:08
136.4 C 225kPa 4:38
136.2 C 221kPa 5:08
136.2 C 223kPa 5:38
136.3 C 224kPa 6:08
136.2 C 223kPa 6:38
136.2 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:29
CYCLE COMPLETE 20:02

Digital Signature #
8D69BAEB104C47A6
