
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000236
7:06 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:21
STERILIZING 5:08
136.0 C 229kPa 5:08
136.3 C 231kPa 5:38
136.4 C 232kPa 6:08
136.4 C 233kPa 6:38
136.1 C 229kPa 7:08
136.4 C 232kPa 7:38
136.3 C 231kPa 8:08
136.4 C 230kPa 8:38
136.4 C 230kPa 8:38
VENTING 8:38
AIR DRYING 9:42
CYCLE COMPLETE 13:30

Digital Signature #
0C7C53F5130A55FD
