

-----  
Midmark M11 - v1.0.0  
Total Cycles: 966

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2021 09:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
2:00 197.7 F 2.8  
4:00 207.5 F 7.2  
6:00 220.4 F 13.0  
8:00 226.9 F 7.8  
10:00 238.5 F 13.0  
12:00 238.3 F 10.4  
14:00 237.7 F 8.5  
16:00 250.1 F 15.4  
18:00 261.8 F 22.0  
20:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.6 F 29.1  
2:00 271.7 F 28.3  
3:00 272.7 F 29.1  
4:00 271.7 F 28.3  
  
Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 223.0 F 0.0  
10:00 231.5 F 0.0  
15:00 232.1 F 0.0  
20:00 226.4 F 0.0  
25:00 223.5 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 20:19  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

07/06/2021 10:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01