

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001773  
9:25 15/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:52  
STERILIZING 3:50  
136.4 C 220kPa 3:50  
135.9 C 221kPa 4:20  
136.4 C 223kPa 4:50  
136.4 C 219kPa 5:20  
136.3 C 222kPa 5:50  
136.4 C 224kPa 6:20  
136.4 C 224kPa 6:50  
136.3 C 223kPa 7:20  
136.3 C 223kPa 7:20  
VENTING 7:20  
AIR DRYING 8:09  
CYCLE COMPLETE 38:09

-----  
Digital Signature #  
19A9602E800D4916  
-----