
Midmark M11 - v1.0.0
Total Cycles: 1204

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2021 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 150.4 F 1.1
4:00 163.5 F 4.2
6:00 188.8 F 8.5
8:00 208.5 F 8.9
10:00 222.4 F 7.5
12:00 234.3 F 12.3
14:00 235.0 F 9.3
16:00 237.5 F 9.4
18:00 246.7 F 13.8
20:00 258.9 F 20.5
22:00 269.8 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.5 F 29.2
2:00 271.8 F 28.8
3:00 272.5 F 29.1
4:00 271.6 F 28.7

Min 271.6 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 214.7 F 0.0
10:00 199.8 F 0.0
15:00 202.6 F 0.0
20:00 196.5 F 0.0
25:00 195.4 F 0.0
30:00 186.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:56
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

09/27/2021 09:08 AM
ACCEPTED: X
REJECTED:
Operator: U01