
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001637
19:16 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 3:55
136.4 C 225kPa 3:55
136.1 C 223kPa 4:25
136.4 C 225kPa 4:55
136.1 C 221kPa 5:25
136.1 C 223kPa 5:55
136.1 C 224kPa 6:25
136.1 C 222kPa 6:55
136.4 C 226kPa 7:25
136.4 C 226kPa 7:25
VENTING 7:25
AIR DRYING 8:21
CYCLE COMPLETE 8:48

Digital Signature #
FB00435CDF7CB979
