

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3797

Name: RITTER2  
Sterilizer ID: V2313228  
07/18/2024 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 89.2 F 0.0  
2:00 90.2 F 0.0  
4:00 104.8 F 0.1  
6:00 134.4 F 1.6  
8:00 163.4 F 5.2  
10:00 189.1 F 9.2  
12:00 210.7 F 13.1  
14:00 219.6 F 7.5  
16:00 232.5 F 12.1  
18:00 233.1 F 9.0  
20:00 238.8 F 10.2  
22:00 246.0 F 13.8  
24:00 257.9 F 20.3  
26:00 268.8 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.8 F 29.9  
2:00 271.5 F 29.1  
3:00 272.7 F 30.0  
4:00 272.0 F 29.3

Min 271.5 F 29.0  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 186.9 F 0.0  
10:00 213.4 F 0.0  
15:00 212.2 F 0.0  
20:00 208.8 F 0.0  
25:00 209.6 F 0.0  
30:00 212.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:10  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:05:28  
CYCLE COMPLETE

07/18/2024 12:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_