

-----  
Midmark M11 - v1.0.5  
Total Cycles: 602

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2021 10:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.7 F 0.0  
2:00 122.9 F 1.2  
4:00 134.5 F 3.6  
6:00 162.5 F 7.2  
8:00 188.8 F 11.7  
10:00 213.4 F 4.2  
12:00 224.4 F 8.8  
14:00 237.9 F 12.3  
16:00 237.9 F 10.8  
18:00 235.5 F 9.0  
20:00 249.2 F 15.5  
22:00 260.7 F 22.0  
24:00 271.1 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.4 F 29.4  
2:00 272.9 F 29.8  
3:00 272.0 F 29.6  
4:00 272.0 F 29.3

Min 271.6 F 29.1  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.0 F 0.0  
5:00 228.4 F 0.0  
10:00 223.6 F 0.0  
15:00 205.5 F 0.0  
20:00 211.8 F 0.0  
25:00 214.0 F 0.0  
30:00 215.2 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 24:39  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:02:54  
CYCLE COMPLETE

12/06/2021 11:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_