
Midmark M11 - v1.0.5
Total Cycles: 942

Name: RITTER2
Sterilizer ID: V2313228
04/01/2022 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.3 F 0.0
2:00 192.9 F 2.6
4:00 207.9 F 8.6
6:00 226.4 F 9.2
8:00 233.1 F 11.5
10:00 236.1 F 10.0
12:00 236.4 F 9.4
14:00 251.6 F 16.8
16:00 263.8 F 24.0
18:00 271.3 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.0
2:00 271.9 F 29.3
3:00 271.8 F 29.8
4:00 272.6 F 29.9

Min 271.3 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 227.6 F 0.0
10:00 220.7 F 0.0
15:00 219.0 F 0.0
20:00 216.9 F 0.0
25:00 222.8 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 2:32
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

04/01/2022 09:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____